



America

# CERTIFICATE

No. U8 094770 0007 Rev. 00

**Holder of Certificate:** Tornado Spectral Systems Inc.  
5155 Spectrum Way, Unit 6  
Mississauga ON L4W 5A1  
CANADA

**Certification Mark:**



**Product:** Laboratory equipment  
**Brand Name:** Raman Spectroscopy  
**Model(s):** tornado spectral systems  
**SuperFlux 785**

**Parameters:** Rated Input Voltage: 100-240 V~  
Rated Frequency: 50/60 Hz  
Rated Input Current: 1.1 A max  
Protection Class: I

**Tested according to:** CSA C22.2 No. 61010-1:2012/U3:2023-06  
UL 61010-1:2012/R:2023-06

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

**Test report no.:** 7169013587-000

**Date,** 2023-12-04

( Peter Keith )